



Sea-Bird Electronics, Inc.

13431 NE 20th St. Bellevue, Washington 98005 USA
www.seabird.com

Phone: (425) 643-9866

Fax: (425) 643-9954

Email: seabird@seabird.com

Pressure Test Certificate

Test Date: **2017-03-28**

Description: **SUNA-V2**

Sensor Information:

Model Number: **SUNA-V2**

Serial Number: **1009**

Pressure Test Protocol:

Low Pressure Test: **40** PSI Held For: **15** Minutes

High Pressure Test: **160** PSI Held For: **15** Minutes

Passed Test: **True**

Tested By: **bk**

**High pressure is
generally equal
to the maximum
depth rating of
the instrument**

