



**Sea-Bird Electronics, Inc.**

13431 NE 20<sup>th</sup> St. Bellevue, Washington 98005 USA  
www.seabird.com

Phone: (425) 643-9866

Fax: (425) 643-9954

Email: seabird@seabird.com

## Pressure Test Certificate

Test Date: **2017-03-31**

Description: **SBE-19Plus SeaCat Profiler**

### Sensor Information:

Model Number: **SBE-19P**

Serial Number: **7798**

### Pressure Test Protocol:

Low Pressure Test: **40**      PSI      Held For: **15**      Minutes

High Pressure Test: **870**      PSI      Held For: **15**      Minutes

Passed Test: **True**

Tested By: **AM**

**High pressure is  
generally equal  
to the maximum  
depth rating of  
the instrument**

