

# CALIBRATION CERTIFICATE

NAME : COMPACT-CT

MODEL : ACT-HR

SERIAL No. : 1301

Parameter : Temperature  
Conductivity



JFE Advantech Co., Ltd.

# Temperature Calibration Certificate

Model : ACT-HR  
 Serial No. : 1301  
 Date : December 01, 2015  
 Location : Production Section  
 Method : Calibration equation is determined from third order regression of samples of the reference temperature against A/D values. Samples are taken at approximately 3, 10, 17, 24, and 31 °C.

1. Equation Instrument temperature[°C] =  $A+B \times N+C \times N^2+D \times N^3$  N: A/D value

2. Coefficients  
 A = -8.057364e00  
 B = 1.074636e-03  
 C = -8.430444e-09  
 D = 9.245523e-14

3. Calibration results

Reference temperature [°C]	A/D value	Instrument temperature [°C]	Residual error [°C]	Acceptance [°C]	OK/NG
2.658	10775	2.659	0.001	±0.050	OK
9.862	18892	9.859	-0.003	±0.050	OK
16.647	27024	16.652	0.005	±0.050	OK
23.662	35566	23.659	-0.003	±0.050	OK
30.398	43560	30.399	0.001	±0.050	OK

4. Verification

Criteria of judgement : Residual error of the instrument temperature at arbitrary point is within the acceptance value.

Reference temperature [°C]	Instrument temperature [°C]	Residual error [°C]	Acceptance [°C]	Judgement
19.715	19.715	0.000	±0.050	Passed

Examined M. Kano

Approved A. Fukuoaka

# Conductivity Calibration Certificate

Model : ACT-HR  
 Serial No. : 1301  
 Date : December 01, 2015  
 Location : Production Section  
 Method : Calibration equation is determined from linear regression of samples of the reference conductivity against A/D values. Samples are taken at approximately 20, 30, 40, and 50 mS/cm.

1. Equation Instrument conductivity[mS/cm] = A+B × N N: A/D value

2. Coefficients A = -6.628440e-01 B = 1.010598e-03

3. Calibration results

Reference conductivity [mS/cm]	A/D value	Instrument conductivity [mS/cm]	Residual error [mS/cm]	Acceptance [mS/cm]	OK/NG
19.409	19862	19.410	0.001	±0.050	OK
30.456	30792	30.455	-0.001	±0.050	OK
39.981	40216	39.979	-0.002	±0.050	OK
51.492	51609	51.493	0.001	±0.050	OK

4. Verification

Criteria of judgement : Residual error of the instrument conductivity at arbitrary point is within the acceptance value.

Reference conductivity [mS/cm]	Instrument conductivity [mS/cm]	Residual error [mS/cm]	Acceptance [mS/cm]	Judgement
45.752	45.750	-0.002	±0.050	Passed

Examined M. Kano  
 Approved A. FukuoKa