

Metadata

Standardized Instrument Name	Idronaut Ocean Seven 304 CTD
Scheme URI	http://vocab.nerc.ac.uk/collection/L22/current/TOOL0861/
Term URI	SDN:L22::TOOL0861
Instrument DOI	
Instrument Unique ID	82/EQ
ID Type	0306147
Model No.	304
Instrument Type	Baseline
Description	MBGL ID 82/EQ
Manufacturer	Idronaut
Manufacturer type	Organizational
Notes	Idronaut 0306147 was calibrated for Freshwater.
Sensor Details	
Sensor Details 1	
Instrument Sensor Name	Ocean Seven 304 CTD Pressure Sensor
Instrument Sensor Serial No.	0306147
Sensor Range	0 to 1000
Sensor Sensitivity	0.05%
Sensor Units	dBar
Last Calibration Date	2019-01-15
Sensor Details 2	
Instrument Sensor Name	Ocean Seven 304 CTD Temperature Sensor
Instrument Sensor Serial No.	0306147
Sensor Range	-5 to +35
Sensor Sensitivity	0.005
Sensor Units	°C
Last Calibration Date	2019-01-15
Sensor Details 3	
Instrument Sensor Name	Ocean Seven 304 CTD Conductivity Sensor

Instrument Sensor Serial No.	0306147
Sensor Range	0 to 6400
Sensor Sensitivity	1
Sensor Units	µS/cm
Last Calibration Date	2019-01-15

Sensor Details 4

Instrument Sensor Name	Ocean Seven 304 CTD Turbidity Sensor
Instrument Sensor Serial No.	0306147
Sensor Range	25, 125, 500, >750
Sensor Sensitivity	200, 40, 10, 2 mV/FTU
Sensor Units	FTU
Last Calibration Date	2019-01-15

Related Datasets

Title Lake Winnipegosis CTD Idronaut data

URL <https://canwin-datahub.ad.umanitoba.ca/data/fr/dataset/mbgl-lwo-idronaut>

Title Waterhen Lake CTD Idronaut Data

URL <https://canwin-datahub.ad.umanitoba.ca/data/fr/dataset/mbgl-lwh-idronaut-data>